



**University of
Zurich**^{UZH}

**Zurich Open Repository and
Archive**

University of Zurich
University Library
Strickhofstrasse 39
CH-8057 Zurich
www.zora.uzh.ch

Year: 2018

Comparison of tissue oximeters on a liquid phantom with adjustable optical properties: an extension

Kleiser, S ; Ostojic, D ; Andresen, B ; Nasseri, N ; Isler, H ; Scholkmann, F ; Karen, T ; Greisen, G ;
Wolf, M

Abstract: Cerebral near-infrared spectroscopy (NIRS) oximetry may help clinicians to improve patient treatment. However, the application of NIRS oximeters is increasingly causing confusion to the users due to the inconsistency of tissue oxygen haemoglobin saturation (StO₂) readings provided by different oximeters. To establish a comparability of oximeters, in our study we performed simultaneous measurements on the liquid phantom mimicking properties of neonatal heads and compared the tested device to a reference NIRS oximeter (OxiplexTS). We evaluated the NIRS oximeters FORE-SIGHT, NIRO and SenSmart, and reproduced previous results with the INVOS and OxyPrem v1.3 oximeters. In general, linear relationships of the StO₂ values with respect to the reference were obtained. Device specific hypoxic and hyperoxic thresholds (as used in the SafeBoosC study, www.safeboosc.eu) and a table allowing for conversion of StO₂ values are provided.

DOI: <https://doi.org/10.1364/BOE.9.000086>

Posted at the Zurich Open Repository and Archive, University of Zurich

ZORA URL: <https://doi.org/10.5167/uzh-144744>

Journal Article

Published Version

Originally published at:

Kleiser, S; Ostojic, D; Andresen, B; Nasseri, N; Isler, H; Scholkmann, F; Karen, T; Greisen, G; Wolf, M (2018). Comparison of tissue oximeters on a liquid phantom with adjustable optical properties: an extension. *Biomedical Optics Express*, 9(1):86-101.

DOI: <https://doi.org/10.1364/BOE.9.000086>



Comparison of tissue oximeters on a liquid phantom with adjustable optical properties: an extension

S. KLEISER,¹ D. OSTOJIC,¹ B. ANDRESEN,² N. NASSERI,^{1,3}
H. ISLER,¹ F. SCHOLKMANN,^{1,3} T. KAREN,⁴ G. GREISEN,² AND
M. WOLF^{1,*}

¹Biomedical Optics Research Laboratory, Department of Neonatology, University Hospital Zurich, University of Zurich, Zurich, Switzerland

²Department of Neonatology, Copenhagen University Hospital, Rigshospitalet, Copenhagen, Denmark

³Institute of Complementary Medicine, University of Bern, Bern, Switzerland

⁴Department of Neonatology, University Hospital Zurich, Zurich, Switzerland

*martin.wolf@usz.ch

Abstract: Cerebral near-infrared spectroscopy (NIRS) oximetry may help clinicians to improve patient treatment. However, the application of NIRS oximeters is increasingly causing confusion to the users due to the inconsistency of tissue oxygen haemoglobin saturation (StO₂) readings provided by different oximeters. To establish a comparability of oximeters, in our study we performed simultaneous measurements on the liquid phantom mimicking properties of neonatal heads and compared the tested device to a reference NIRS oximeter (OxiplexTS). We evaluated the NIRS oximeters FORE-SIGHT, NIRO and SenSmart, and reproduced previous results with the INVOS and OxyPrem v1.3 oximeters. In general, linear relationships of the StO₂ values with respect to the reference were obtained. Device specific hypoxic and hyperoxic thresholds (as used in the SafeBoosC study, www.safeboosc.eu) and a table allowing for conversion of StO₂ values are provided.

© 2017 Optical Society of America under the terms of the [OSA Open Access Publishing Agreement](#)

OCIS codes: (120.3890) Medical optics instrumentation; (170.6510) Spectroscopy, tissue diagnostics.

References and links

1. S. Hyttel-Sorensen, A. Pellicer, T. Alderliesten, T. Austin, F. van Bel, M. Benders, O. Claris, E. Dempsey, A. R. Franz, M. Fumagalli, C. Gluud, B. Grevstad, C. Hagmann, P. Lemmers, W. van Oeveren, G. Pichler, A. M. Plomgaard, J. Riera, L. Sanchez, P. Winkel, M. Wolf, and G. Greisen, "Cerebral near infrared spectroscopy oximetry in extremely preterm infants: phase ii randomised clinical trial," *BMJ (Clinical research ed.)* **350**, g7635 (2015).
2. P. E. Bickler, J. R. Feiner, and M. D. Rollins, "Factors affecting the performance of 5 cerebral oximeters during hypoxia in healthy volunteers," *Anesth. Analg.* **117**, 813–823 (2013).
3. S. Hyttel-Sorensen, T. W. Hessel, and G. Greisen, "Peripheral tissue oximetry: comparing three commercial near-infrared spectroscopy oximeters on the forearm," *J. Clin. Monit. Comput.* **28**, 149–155 (2014).
4. T. W. Hessel, S. Hyttel-Sorensen, and G. Greisen, "Cerebral oxygenation after birth - a comparison of invos and fore-sight near-infrared spectroscopy oximeters," *Acta Paediatr.* **103**, 488–493 (2014).
5. L. M. Dix, F. van Bel, W. Baerts, and P. M. Lemmers, "Comparing near-infrared spectroscopy devices and their sensors for monitoring regional cerebral oxygen saturation in the neonate," *Pediatr. Res.* **74**, 557–563 (2013).
6. S. Hyttel-Sorensen, L. C. Sorensen, J. Riera, and G. Greisen, "Tissue oximetry: a comparison of mean values of regional tissue saturation, reproducibility and dynamic range of four nirs-instruments on the human forearm," *Biomed. Opt. Express* **2**, 3047–3057 (2011).
7. A. Schneider, B. Minnich, E. Hofstätter, C. Weisser, E. Hattinger-Jürgenssen, and M. Wald, "Comparison of four near-infrared spectroscopy devices shows that they are only suitable for monitoring cerebral oxygenation trends in preterm infants," *Acta Paediatr.* **103**, 934–938 (2014).
8. K. L. Tomlin, A.-M. Neitenbach, and U. Borg, "Detection of critical cerebral desaturation thresholds by three regional oximeters during hypoxia: a pilot study in healthy volunteers," *BMC Anesthesiol.* **17**, 6 (2017).
9. A. Dullenkopf, B. Frey, O. Baenziger, A. Gerber, and M. Weiss, "Measurement of cerebral oxygenation state in anaesthetized children using the invos 5100 cerebral oximeter," *Paediatr. Anaesth.* **13**, 384–391 (2003).
10. L. C. Sorensen and G. Greisen, "Precision of measurement of cerebral tissue oxygenation index using near-infrared spectroscopy in preterm neonates," *J. Biomed. Opt.* **11**, 054005 (2006).

11. L. M. L. Dix, F. van Bel, and P. M. A. Lemmers, "Monitoring cerebral oxygenation in neonates: An update," *Front. Pediatr.* **5** (2017).
12. W. M. Coplin, G. E. O'Keefe, M. S. Grady, G. A. Grant, K. S. March, H. R. Winn, and A. M. Lam, "Thrombotic, infectious, and procedural complications of the jugular bulb catheter in the intensive care unit," *Neurosurgery* **41**, 101–107 (1997).
13. B. Meyer, C. Schaller, C. Frenkel, B. Ebeling, and J. Schramm, "Distributions of local oxygen saturation and its response to changes of mean arterial blood pressure in the cerebral cortex adjacent to arteriovenous malformations," *Stroke* **30**, 2623–2630 (1999).
14. H. D. Clay, "Validity and reliability of the SjO_2 catheter in neurologically impaired patients: a critical review of the literature," *J. Neurosci. Nurs.* **32**, 194–203 (2000).
15. H. Ito, I. Kanno, H. Iida, J. Hatazawa, E. Shimosegawa, H. Tamura, and T. Okudera, "Arterial fraction of cerebral blood volume in humans measured by positron emission tomography," *Ann. Nucl. Med.* **15**, 111–116 (2001).
16. N. C. Brun, A. Moen, K. Borch, O. D. Saugstad, and G. Greisen, "Near-infrared monitoring of cerebral tissue oxygen saturation and blood volume in newborn piglets," *Am. J. Physiol.* **273**, H682–H686 (1997).
17. P. B. Benni, B. Chen, F. D. Dykes, S. F. Wagoner, M. Heard, A. J. Tanner, T. L. Young, K. Rais-Bahrami, O. Rivera, and B. L. Short, "Validation of the cas neonatal nirs system by monitoring vv-ecmo patients: preliminary results," *Adv. Exp. Med. Biol.* **566**, 195–201 (2005).
18. H. M. Watzman, C. D. Kurth, L. M. Montenegro, J. Rome, J. M. Steven, and S. C. Nicolson, "Arterial and venous contributions to near-infrared cerebral oximetry," *Anesthesiology* **93**, 947–953 (2000).
19. M. Wolf, G. Naulaers, F. van Bel, S. Kleiser, and G. Greisen, "A review of near-infrared spectroscopy for term and preterm newborns," *J. Near. Infrared Spectrosc.* **20**, 43–55 (2012).
20. P. B. Benni, D. MacLeod, K. Ikeda, and H.-M. Lin, "A validation method for near-infrared spectroscopy based tissue oximeters for cerebral and somatic tissue oxygen saturation measurements," *J. Clin. Monit. Comput.* (2017).
21. H. Sorensen, N. H. Secher, and P. Rasmussen, "A note on arterial to venous oxygen saturation as reference for nirs-determined frontal lobe oxygen saturation in healthy humans," *Front. Physiol.* **4**, 403 (2013).
22. P. E. Bickler, J. R. Feiner, M. S. Lipnick, P. Batchelder, D. B. MacLeod, and J. W. Severinghaus, "Effects of acute, profound hypoxia on healthy humans: Implications for safety of tests evaluating pulse oximetry or tissue oximetry performance," *Anesth. Analg.* **124**, 146–153 (2017).
23. S. N. Davie and H. P. Grocott, "Impact of extracranial contamination on regional cerebral oxygen saturation: a comparison of three cerebral oximetry technologies," *Anesthesiology* **116**, 834–840 (2012).
24. H. Sørensen, P. Rasmussen, C. Siebenmann, M. Zaar, M. Hvidtfeldt, S. Ogoh, K. Sato, M. Kohl-Bareis, N. H. Secher, and C. Lundby, "Extra-cerebral oxygenation influence on near-infrared-spectroscopy-determined frontal lobe oxygenation in healthy volunteers: a comparison between invos-4100 and niro-200nx," *Clin. Physiol. Funct. Imaging.* **35**, 177–184 (2015).
25. M. Dehaes, P. E. Grant, D. D. Sliva, N. Roche-Labarbe, R. Pienaar, D. A. Boas, M. A. Franceschini, and J. Selb, "Assessment of the frequency-domain multi-distance method to evaluate the brain optical properties: Monte carlo simulations from neonate to adult," *Biomed. Opt. Express* **2**, 552–567 (2011).
26. M. A. Franceschini, S. Fantini, L. A. Paunescu, J. S. Maier, and E. Gratton, "Influence of a superficial layer in the quantitative spectroscopic study of strongly scattering media," *Appl. Opt.* **37**, 7447–7458 (1998).
27. M. Firbank and D. T. Delpy, "A design for a stable and reproducible phantom for use in near-infrared imaging and spectroscopy," *Phys. Med. Biol.* **38**, 847–853 (1993).
28. L. A. Dempsey, M. Persad, S. Powell, D. Chitnis, and J. C. Hebden, "Geometrically complex 3d-printed phantoms for diffuse optical imaging," *Biomed. Opt. Express* **8**, 1754–1762 (2017).
29. P. Diep, S. Pannem, J. Sweer, J. Lo, M. S. G. Stueber, Y. Zhao, S. Tabassum, R. Istfan, J. Wu, S. Erramilli, and D. Roblyer, "Three-dimensional printed optical phantoms with customized absorption and scattering properties," *Biomed. Opt. Express* **6**, 4212–4220 (2015).
30. S. Suzuki, S. Takasaki, T. Ozaki, and Y. Kobayashi, "A tissue oxygenation monitor using nir spatially resolved spectroscopy," *SPIE* **3597**, 582–592 (1999).
31. J. G. Kim and H. Liu, "Variation of haemoglobin extinction coefficients can cause errors in the determination of haemoglobin concentration measured by near-infrared spectroscopy," *Phys. Med. Biol.* **52**, 6295–6322 (2007).
32. E. L. Hull, M. G. Nichols, and T. H. Foster, "Quantitative broadband near-infrared spectroscopy of tissue-simulating phantoms containing erythrocytes," *Phys. Med. Biol.* **43**, 3381–3404 (1998).
33. A. Bozkurt, A. Rosen, H. Rosen, and B. Onaral, "A portable near-infrared spectroscopy system for bedside monitoring of newborn brain," *Biomed. Eng. Online* **4**, 29 (2005).
34. C. D. Kurth, H. Liu, W. S. Thayer, and B. Chance, "A dynamic phantom brain model for near-infrared spectroscopy," *Phys. Med. Biol.* **40**, 2079–2092 (1995).
35. X. Lv, Y. Xue, H. Wang, S. W. Shen, X. Zhou, G. Liu, E. Dong, and R. X. Xu, "3d printing of microtube in solid phantom to simulate tissue oxygenation and perfusion (conference presentation)," *SPIE* **10056**, 1005606–1 (2017).
36. E. N. Marieb and K. N. Hoehn, *Human Anatomy and Physiology* (Pearson, 2000).
37. H. Liu, B. Chance, A. Hielscher, S. Jacques *et al.*, "Influence of blood vessels on the measurement of hemoglobin oxygenation as determined by time-resolved reflectance spectroscopy," *Med. Phys.* **22**, 1209–1217 (1995).
38. S. Kleiser, N. Nasser, B. Andresen, G. Greisen, and M. Wolf, "Comparison of tissue oximeters on a liquid phantom with adjustable optical properties," *Biomed. Opt. Express* **7**, 2973–2992 (2016).

39. M. A. Franceschini, S. Fantini, A. E. Cerussi, B. B. Barbieri, B. Chance, and E. Gratton, "Quantitative spectroscopic determination of hemoglobin concentration and saturation in a turbid medium: analysis of the effect of water absorption," *J. Biomed. Opt.* **2**, 147–153 (1997).
40. S. Hyttel-Sorensen, T. W. Hessel, A. la Cour, and G. Greisen, "A comparison between two nirs oximeters (invos, oxyprem) using measurement on the arm of adults and head of infants after caesarean section," *Biomed. Opt. Express* **5**, 3671–3683 (2014).
41. S. Kleiser, S. Hyttel-Sorensen, G. Greisen, and M. Wolf, "Comparison of near-infrared oximeters in a liquid optical phantom with varying intralipid and blood content," *Adv. Exp. Med. Biol.* **876**, 413–418 (2016).
42. N. Nasser, S. Kleiser, D. Ostojic, T. Karen, and M. Wolf, "Quantifying the effect of adipose tissue in muscle oximetry by near infrared spectroscopy," *Biomed. Opt. Express* **7**, 4605–4619 (2016).
43. D. MacLeod, K. Ikeda, C. Cheng, and C. Shaw, "Validation of the next generation fore-sight elite tissue oximeters for adult cerebral tissue oxygen saturation," *Anesth. Analg.* **116**, 1–182 (2013).
44. Nonin, "Model 8004CA, Instructions for use - english," Tech. Rep. 9539-001-03, Nonin Medical, Inc., 13700 1st Avenue North Plymouth, MN, 55441-5443, USA (2016).
45. Nonin, "Model 8004CB-NA, Instructions for use - english," Tech. Rep. 9223-001-03, Nonin Medical, Inc., 13700 1st Avenue North Plymouth, MN, 55441-5443, USA (2016).
46. R. N. Kreeger, C. Ramamoorthy, S. C. Nicolson, W. A. Ames, R. Hirsch, L. F. Peng, A. C. Glatz, K. D. Hill, J. Hoffman, J. Tomasson, and C. D. Kurth, "Evaluation of pediatric near-infrared cerebral oximeter for cardiac disease," *Ann. Thorac. Surg.* **94**, 1527–1533 (2012).
47. C. Benkwitz, M. C. Hamilton, D. R. Janssen, and T. P. Doyle, "Validation of the fore-sight elite tissue oximeter in context with simultaneous vital sign recording in pediatric patients undergoing cardiac catheterization," *J. Thorac. Cardiovasc. Surg.* **146**, 1153 (2013).
48. D. M. Hueber, S. Fantini, A. E. Cerussi, and B. Barbieri, "New optical probe designs for absolute (self-calibrating) nir tissue hemoglobin measurements," *SPIE* **3597**, 618–631 (1999).
49. S. J. Arri, T. Muehleman, M. Biallas, H. U. Bucher, and M. Wolf, "Precision of cerebral oxygenation and hemoglobin concentration measurements in neonates measured by near-infrared spectroscopy," *J. Biomed. Opt.* **16**, 047005 (2011).
50. J. Dobbing and J. Sands, "Quantitative growth and development of human brain," *Arch. Dis. Child.* **48**, 757–767 (1973).
51. K. Linnet, "Evaluation of regression procedures for methods comparison studies," *Clin. Chem.* **39**, 424–432 (1993).
52. W. G. Zijlstra, A. Buursma, and W. P. Meeuwse-van der Roest, "Absorption spectra of human fetal and adult oxyhemoglobin, de-oxyhemoglobin, carboxyhemoglobin, and methemoglobin," *Clin. Chem.* **37**, 1633–1638 (1991).
53. Y. A. Wickramasinghe, K. S. Palmer, R. Houston, S. A. Spencer, P. Rolfe, M. S. Thorniley, B. Oeseburg, and W. Colier, "Effect of fetal hemoglobin on the determination of neonatal cerebral oxygenation by near-infrared spectroscopy," *Pediatr. Res.* **34**, 15–17 (1993).
54. I. Uchida, C. Tashiro, Y. H. Koo, T. Mashimo, and I. Yoshiya, "Carboxyhemoglobin and methemoglobin levels in banked blood," *J. Clin. Anesth.* **2**, 86–90 (1990).

1. Introduction

Near-infrared spectroscopy (NIRS) is a technique for measuring oxygenation of tissue non-invasively and continuously. One important application may be to prevent cerebral haemorrhagic and ischaemic insults in preterm infants. A randomized controlled trial safeguarding the brains of our smallest children (SafeBoosC) showed that it is possible to reduce the cerebral hypoxic/hyperoxic burden in extremely preterm infants, when combining NIRS monitoring with a treatment guideline [1]. This also led to a substantially reduced mortality and incidence of severe brain lesions in the treatment group. However, this reduction was not statistically significant and therefore a large study is planned to confirm the clinical benefits. One major issue here is that different brands of NIRS oximeters provide systematically different oxygenation values, as shown by numerous studies [2–8]. Currently, tissue oxygen haemoglobin saturation (StO_2) values cannot be compared between different oximeters and sensors [9, 10]. This constitutes a problem when setting alarm limits in the mentioned new trial or when interpreting the literature [11].

Why do different NIRS oximeters provide different values? NIRS oximeters measure absolute values of the StO_2 , which represents the proportion of oxygenated haemoglobin of all haemoglobin present in arterial, capillary and venous compartments of the tissue interrogated by the sensor. The reason for the different values between different brands is that it is difficult to validate. One StO_2 *in vivo* approach is to take arterial and venous (jugular vein for cerebral measurements) blood samples and to determine their oxygen haemoglobin saturation (SO_2) by co-oximetry, a

trusted method.

An ethical problem here is that venous blood samples from the jugular bulb cannot be collected without risk to the patient [12]. In particular in preterm infants, sampling the jugular bulb is clinically not feasible. For ethical reasons, these studies are therefore typically either performed in healthy young adults or patients requiring a jugular bulb catheter for clinical reasons. For the latter an influence of their pathology is likely. Both scenarios may not be considered fully representative of the wide range of patients in hospitals.

A methodological problem is that the venous jugular bulb blood represents an average of the whole brain hemisphere, which is much larger than the volume of the NIRS measurement it is being compared to [13]. In addition, there may be extra-cerebral contamination [14]. To determine the StO_2 the arterial-to-venous volume ratio (AVR) has to be assumed (typically 25:75 % or 30:70 %). Although a positron emission tomography (PET) study estimated AVR to be 30:70 [15], this is a source of errors, because the AVR was shown to vary considerably between individuals [16–18]. Furthermore, the AVR depends on the specific location measured and it changes over time [19], in particular during desaturation experiments [2]. The AVR is further influenced by e.g. the end tidal CO_2 , which must be kept constant [20] and the subject's positioning [2] which may cause systematic differences between studies. A change in the group AVR assumption by 10 % (i.e. 20:80 % or 40:60 %) causes a change in bias of already ± 3 % [20]. But a recent meta-analysis on published cerebral StO_2 acquired with the INVOS oximeter reported that even higher AVR of up to 75:25 % fit their data best [21]. These examples show that already the *in vivo* reference StO_2 is associated with a number of assumptions and uncertainties on group and patient levels, is subject to considerable random and systematic errors, and does not constitute a 'gold standard'.

Apart from problems with the reference StO_2 , there are further problems encountered during *in vivo* validation by desaturation studies: Although short episodes of hypoxia with arterial oxygen haemoglobin saturation (SaO_2) as low as 50-70 % are tolerated well by healthy adults [22], the very low range of StO_2 cannot be validated safely with this method, leaving questions on the validity of extreme (high and low) readings which may occur in clinical situations.

Another issue is that changes in StO_2 originate from the brain and partly from extra-cerebral layers. In adults a significant influence of scalp and skin on the cerebral StO_2 values was found [23,24] even though the effect of the skull layer, which is likely substantial [25,26], was neglected. However, StO_2 is changed everywhere and not only in the brain in desaturation studies. Thus, the ability of a NIRS oximeter to measure brain StO_2 independently of the more superficial tissues is not assessed by current methods.

A desaturation study comparing several cerebral oximeters found 'significantly more positive bias at lower SaO_2 ' in some oximeters [2], which corresponds to lower sensitivity to oxygenation changes. The difference between oximeters is striking, because most oximeters in that study were calibrated by this method of arterial and venous blood sampling. Hence, this approach results in inconsistent StO_2 values between different instruments, and therefore seems inappropriate for validation of cerebral StO_2 .

A further option for *in vivo* validation is the vascular occlusion test on extremities. The ischaemia induced by the occlusion allows a wide range of StO_2 , but otherwise the same mentioned issues as for the brain remain unresolved.

Phantoms, on the other hand, have the advantage of controllable optical properties and can be adjusted for specific research questions and may include sophisticated geometries. In 1993, Firbank et. al. recommended a solid phantom made of optically clear polyester resin. Dyes and titanium dioxide were added to adjust its optical properties [27]. 3D printers allow producing anatomically accurate, tissue-equivalent phantoms of infant heads [28] and mice [29]. However, since oximeters include a different set of wavelengths each, it is more appropriate to include real hemoglobin to obtain an accurate absorption spectrum. In addition, StO_2 and total haemoglobin

concentration (c_{tHb}) can be changed over the entire relevant range [30–33]. One example is the "dynamic phantom brain model" [34]. This model consisted of (1) a resin with a vascular network (500 μm diameter) which represents brain and is perfused by human blood, (2) a bubble oxygenator to change the oxygenation of the blood and (3) a roller pump [34]. Another method is to fabricate microtubes for the vascular network [35]. The diameter of these microtubes, however, was still larger (300 μm inner diameter) than the one of human capillaries (9 μm (inner diameter)) [36] which is a relevant difference for NIRS oximetry [37].

The most promising method to compare NIRS devices in our opinion is a dynamic 2-layer phantom with adjustable optical properties mimicking the neonatal head as described in detail in [38]. This phantom contains Intralipid to adjust scattering and real human haemoglobin. The aim was to apply this phantom to quantitatively compare different NIRS devices for SafeBoosC to establish comparable intervention thresholds between devices. We provided mathematical equations to convert StO_2 between devices. We demonstrated that the systematic differences in the StO_2 values between NIRS oximeters also depend on the c_{tHb} of the phantom, which may be due to different assumptions regarding background absorbers [39].

In our previous comparison studies in phantoms [38,40–42] we observed substantial systematic differences between *in vivo* calibrated NIRS oximeters. Such differences were also shown *in vivo* [2–8] and are probably due to the error prone *in vivo* calibration rather than the simple nature of the phantom [20]. The oximeters included in [38] and here can be assumed to apply the multi-distance approach for calculation of StO_2 , since they provide more than one source-detector separation. Such oximeters are minimally influenced by a superficial layer of 2.5 mm [26,42] independent of epidermal pigmentation which leads to a lower intensity of detected light, thus lowering the SNR. Therefore, our phantom model with a static superficial layer is an appropriate model for neonates. We included adult sensors in our experiments since adult sensors have been used off-label in neonates. In adults the skull/scalp/skin region is > 1 cm thick and extra-cerebral signals contribute substantially to the StO_2 . Therefore, the results are not to be translated to adults. Although it is currently unclear how a good model of the adult head can be achieved best, we have shown that phantoms like ours have the capability to assess sensitivity of instruments to deeper layers [42]. Phantoms provide for simultaneous measurements by different NIRS oximeters on truly the same sample and over a wide StO_2 range. Even though not yet fully completed, it is possible to include a precise and accurate StO_2 reference for validation of accuracy. Thus, phantoms are a versatile tool to validate NIRS oximeters reproducibly.

In this paper, the aim was to extend our previous work [38] with a new set of devices and sensors to provide clinical researchers and clinicians a means to translate neonatal cerebral StO_2 acquired by their oximeter to the results reported by others, using other oximeters. The three additional oximeters NIRO-200NX (Hamamatsu), FORE-SIGHT Elite (Casmed) and SenSmart-X100 (Nonin) were compared with OxiplexTS (ISS) as a reference. Measurements for the INVOS 5100C (Medtronic) and OxyPrem v1.3 (in-house developed, University Hospital Zurich) oximeters were repeated with the aim to test reproducibility of the liquid phantom method.

2. Material and methods

2.1. NIRS oximeters

Some of the instruments used in [38] and this experiment were calibrated *in vivo*, which only probes a limited range of StO_2 safely [20,43–47]. In neonates, however, StO_2 often lies and is considered outside this calibrated range. We thus report data from the whole range recorded. We solely describe oximeters and sensors tested for the first time with this method. For a description of previously tested devices please refer to [38].

The NIRO-200NX (Hamamatsu Photonics K.K., Hamamatsu, Japan) applies an LED source with three wavelengths (735 nm, 810 nm, 850 nm) in disposable adhesive (NIRO small/ NIRO

large) and re-usable (NIRO small RU/NIRO large RU) sensors with source-detector separation (SDS) of 3 cm (small) and 4 cm (large). Detectors of the disposable sensors and the re-usable probes have different shapes. The device combines modified Beer-Lambert law and spatially resolved spectroscopy for trend and absolute measurements. It is approved for clinical use.

The FORE-SIGHT Elite (CAS Medical Systems, Inc., Branford, CT, USA) applies five wavelengths (690, 730, 770, 810, and 870 nm) to measure absolute StO₂. The large sensor (FORE-SIGHT adult) comprises SDS of 1.5 and 5 cm while the medium sensor (FORE-SIGHT medium) has SDS of 1.3 and 4 cm. Small sensors for neonates use SDS of 1 and 2.5 cm and are available as adhesive (FORE-SIGHT small) and non-adhesive version (FORE-SIGHT small band). The device is approved for clinical use.

The SenSmart-X100 (Nonin Medical, Inc., Plymouth, MN, USA) applies four wavelengths (730, 760, 810 and 870 nm). The sensors have two light sources and two detectors giving four light paths with SDS of 2 and 4 cm in case of the adult sensor EQUANOX Advance 8004CA (SenSmart adult). The device is approved for clinical use.

The device we used as reference here, OxiplexTS (ISS, Champaign, Illinois, USA), employs two modulated light sources (692 nm and 834 nm) measuring absolute tissue absorption and reduced scattering coefficients. A rigid sensor with four SDS (2.5 cm, 3.0 cm, 3.5 cm and 4.0 cm) was applied. In contrast to other tissue oximeters, this device is not approved for clinical use, but has a CE mark for research. As we solely intended to compare instruments and not to assess accuracy, our reference does not have to measure the true StO₂, but only needs to measure reproducibly and independent of c_{tHb} and reduced scattering coefficient (μ'_s). OxiplexTS measures both absorption coefficient (μ'_a) and μ'_s and was validated in several publications such as [26, 39, 48] and *in vivo* a precision of 2.0 % was demonstrated in term-born newborns [49], thus fulfilling our needs.

2.2. Phantom setup

As in the previous systematic *in vitro* comparison of NIRS oximeters [38], the setup consisted of a container offering four windows for simultaneous recording with NIRS sensors. Each window was made of a layer of silicone with tailored optical properties and thickness (2.5 mm) to resemble the scalp and skull of a typical neonate. The container was filled with a liquid containing human haemoglobin (Hb) and Intralipid (IL) which was added to obtain the wanted μ'_s of $\approx 5.5 \text{ cm}^{-1}$. Each phantom preparation covered typical c_{tHb} in neonates (see Table 1). OxiplexTS (ISS, Inc., Champaign, IL, USA) was applied as a reference oximeter. The investigation was performed in three phantoms on three different days.

2.2.1. Phantom no. 1

The FORE-SIGHT Elite with FORE-SIGHT small and FORE-SIGHT small band as well as NIRO small were investigated with one deoxygenation per liquid mixture. Then FORE-SIGHT small, FORE-SIGHT small band and NIRO small were removed, remounted and another three deoxygenations were performed. FORE-SIGHT small reported an 'out of range' error in this period and did not provide data.

2.2.2. Phantom no. 2

Two groups of oximeters were investigated: Group 1 with SenSmart adult, FORE-SIGHT adult and NIRO large; Group 2 with FORE-SIGHT medium, NIRO small RU and NIRO large RU. Groups 1 and 2 were intermittently placed on the phantom. Two deoxygenations at a $c_{\text{tHb}} = 30 \mu\text{M}$ were performed with group 1 before switching to group 2 and performing another one. Then c_{tHb} was increased to $47.5 \mu\text{M}$ and one deoxygenation was first measured with group 2 followed by group 1. After increasing c_{tHb} to $75 \mu\text{M}$, there were first two deoxygenations recorded with group 1 and then one with group 2.

2.2.3. Phantom no. 3

Here, the INVOS 5100C with adult SomaSensor SAFB-SM (INVOS adult), OxyPrem v1.3 and SenSmart adult were employed. For $c_{\text{tHb}} = 30 \mu\text{M}$ and $75 \mu\text{M}$ one deoxygenation each was performed. At $c_{\text{tHb}} = 47.5 \mu\text{M}$ we performed several deoxygenations and removed and mounted the sensors in-between them to test robustness to repositioning. During the second $c_{\text{tHb}} = 47.5 \mu\text{M}$ deoxygenation new sensors for INVOS adult and SenSmart adult were used and afterwards the original ones were reapplied.

2.3. Liquid mixture

Table 1. Amount of blood added and resulting c_{tHb} and htc of the liquid phantoms for each of the three mixtures.

Mixture no.	Phantom no.1				Phantom no.2			Phantom no.3		
	c_{tHb} (μM)	Blood (ml)	c_{tHb} (μM)	htc (%)	Blood (ml)	c_{tHb} (μM)	htc (%)	Blood (ml)	c_{tHb} (μM)	htc (%)
1	≈ 30	24	30.67	0.60	22	28.20	0.55	22	28.20	0.55
2	≈ 47.5	39	49.56	0.97	37	47.16	0.92	37	47.16	0.92
3	≈ 75	61	76.86	1.51	59	74.57	1.46	62	78.28	1.53

The phantom consisted of the same ingredients as in [38], obtained from the same suppliers. Main ingredient was phosphate buffered saline (PBS) (Kreis, $pH = 7.4$) with a volume of 2.5 l to which 74 ml IL (20 %) and defined amounts of blood were added. We added sodium bicarbonate buffer (SBB) ($8.4\% \equiv 1 \text{ mmol/ml}$) initially (15 ml) and each time when adding blood to the phantom (10 ml). We added 3 g of yeast solved in a small amount of SBB and 3 ml glucose (50 %) solution to the phantom to trigger deoxygenation. Each time blood was added, another 3 ml of glucose were added. The phantom was re-oxygenated by adding pure oxygen (O_2) by bubbling.

In [38], we covered the typical range of c_{tHb} in neonates with $c_{\text{tHb}} = 25 \mu\text{M}$, $45 \mu\text{M}$ and $70 \mu\text{M}$. Here we adhered to these values. Table 1 shows the amounts of blood added and the effective c_{tHb} of each phantom. Two erythrocyte concentrates with the same measured $c_{\text{tHb}} \equiv 215 \text{ g/l}$ and $\text{htc} \equiv 65.5 \%$ were employed, one for phantom 1 and one for phantoms 2 and 3.

2.4. Data extraction and analysis

To ensure consistency, data processing methods were the same as in [38]. We applied the same StO_2 limits for fitting to both axes, i.e. to OxiplexTS and the oximeter to be investigated. However, here we set the upper limit to $\text{StO}_2 = 85 \%$ (instead of 94%) because StO_2 of FORE-SIGHT Elite was non-linear above this value for all sensors. The lower limit was kept at $\text{StO}_2 = 16 \%$.

3. Results

Figures 1(a)-1(d) and 2(a)-2(d) show StO_2 of the oximeters investigated for the first time, whereas Fig. 3(a) and 3(b) show the results of two oximeters already included in [38]. We added the data from [38] to the same plots for comparison. Table 2 shows the coefficients for linear fits in the range $16 \leq \text{StO}_2 \leq 85 \%$ and their generally high correlation coefficients (R^2) except for SenSmart adult showing a remarkable curvature. We do not report results of FORE-SIGHT medium due to inconsistency with data obtained in another experiment (not published).

Table 3 shows the hypoxic ($\text{StO}_2 = 55 \%$) and hyperoxic thresholds ($\text{StO}_2 = 85 \%$) defined by the SafeBoosC study based on the INVOS adult sensor. To generate table 3, we took the corresponding StO_2 thresholds of the OxiplexTS of ($\text{StO}_2 = 47.0 \%$ and $\text{StO}_2 = 77.2 \%$) as determined in [38] for data at $c_{\text{tHb}} = 47.5 \mu\text{M}$. We then applied linear functions with the

coefficients listed in table 2 for $x = 47\%$, 77.2% . Table 3 also lists the uncertainty of StO_2 readings determined by subtracting the threshold at $c_{\text{tHb}} = 75\ \mu\text{M}$ from the one at $c_{\text{tHb}} = 30\ \mu\text{M}$.

In [38], we provided a table for converting StO_2 from one oximeter to another for typical neonates for mixture 2 ($c_{\text{tHb}} = 45\ \mu\text{M}$ in [38], $c_{\text{tHb}} = 47.5\ \mu\text{M}$ in present study). In tables 4-6 new data are complemented with reprinted results of [38] *in italic*.

We recorded pH and temperature (T) by repeated measurements in all three phantoms. Overall, pH was in the range $7.13 \leq \text{pH} \leq 7.64$ with a general decrease over time and an increase during bubbling of O_2 . Values of temperature were in the range $35.1\ ^\circ\text{C} \leq T \leq 36.4\ ^\circ\text{C}$ with a variation of $\approx 0.5\ ^\circ\text{C}$ within each single phantom. μ'_s as measured by OxiplexTS decreased by 11 % from baseline the end of phantom 1 (after 210 minutes), by 16 % at the end of phantom 2 (after 275 minutes) and by 9 % at the end of phantom 3 (after 160 minutes).

Table 2. Coefficients for linear transformation $\text{StO}_{2, \text{Device}_x} = a * \text{StO}_{2, \text{OxiplexTS}} + b$.

	$c_{\text{tHb}} = 30\ \mu\text{M}$			$c_{\text{tHb}} = 47.5\ \mu\text{M}$			$c_{\text{tHb}} = 75\ \mu\text{M}$		
	b	a	R^2	b	a	R^2	b	a	R^2
FORE-SIGHT small	51.0	0.35	0.998	43.5	0.47	0.997	35.6	0.57	0.994
FORE-SIGHT small band	52.0	0.37	0.998	44.1	0.48	0.997	37.3	0.57	0.994
NIRO small	41.6	0.45	1.000	32.9	0.60	0.999	28.1	0.66	0.999
NIRO small RU	41.9	0.49	0.999	36.1	0.57	0.999	28.1	0.68	0.999
INVOS adult	18.8	0.83	0.999	5.0	1.04	0.998	-11.3	1.28	0.999
OxyPrem V1.3	24.0	0.69	1.000	19.5	0.74	0.998	13.6	0.83	0.998
SenSmart adult	2.7	0.56	0.983	-12.8	0.85	0.992	-34.1	1.21	0.991
FORE-SIGHT adult	39.8	0.48	0.995	40.1	0.52	0.994	29.6	0.66	0.995
NIRO large	36.4	0.65	1.000	26.7	0.75	1.000	15.2	0.92	0.998
NIRO large RU	36.6	0.61	1.000	29.6	0.69	1.000	21.7	0.81	0.999

Table 3. SafeBoosC intervention thresholds for $c_{\text{tHb}} = 47.5\ \mu\text{M}$ and range of uncertainty due to variation of c_{tHb} in the range of $30\ \mu\text{M}$ to $75\ \mu\text{M}$ (threshold at $c_{\text{tHb}} = 75\ \mu\text{M}$ - threshold at $c_{\text{tHb}} = 30\ \mu\text{M}$). All values are given as StO_2 [%].

	hypoxic threshold	uncertainty range due to c_{tHb}	hyperoxic threshold	uncertainty range due to c_{tHb}
Oxiplex TS	47.0		77.2	
FORE-SIGHT small	65.5	4.9	79.6	-1.8
FORE-SIGHT small band	66.7	5.3	81.2	-0.8
NIRO small	61.0	3.5	79.1	-3.0
NIRO small RU	62.8	4.8	80.0	-1.0
INVOS adult	53.7	9.1	84.9	-4.3
OxyPrem V1.3	54.3	3.6	76.7	-0.7
SenSmart adult	27.3	6.1	53.0	-13.7
FORE-SIGHT adult	64.7	2.0	80.5	-3.4
NIRO large	61.9	8.4	84.6	0.1
NIRO large RU	62.2	5.7	83.2	-0.3




Fig. 1. (a) FORE-SIGHT small with fixation band vs. OxiplexTS for three different c_{tHb} . Data was obtained in phantom 1. The sensor was repositioned after the 1st of 4 deoxygenations at $c_{tHb} = 75 \mu\text{M}$. (b) FORE-SIGHT small adhesive vs. OxiplexTS measured in phantom 1. (c) FORE-SIGHT adult vs. OxiplexTS in phantom 2. The sensor was repositioned after $c_{tHb} = 30 \mu\text{M}$. There were two consecutive deoxygenations at $c_{tHb} = 30 \mu\text{M}$ and $75 \mu\text{M}$. (d) SenSmart adult vs. OxiplexTS measured in phantom 2 and 3. In phantom 2, the sensor was repositioned after $c_{tHb} = 30 \mu\text{M}$. There were two consecutive deoxygenations at $c_{tHb} = 30 \mu\text{M}$ and $75 \mu\text{M}$. In phantom 3, there were 3 deoxygenations with sensor repositioning in between at $c_{tHb} = 47.5 \mu\text{M}$ with the second one being performed with a new sensor.




Fig. 2. (a) NIRO-200NX small reusable vs. OxiplexTS three different c_{tHb} measured in phantom 2. The sensor was repositioned after $c_{\text{tHb}} = 47.5\ \mu\text{M}$. (b) NIRO-200NX small single-use vs. OxiplexTS in phantom 1. The sensor was repositioned after the 1st of 4 deoxygenations at $c_{\text{tHb}} = 75\ \mu\text{M}$. (c) NIRO-200NX large reusable vs. OxiplexTS in phantom 2. The sensor was repositioned after $c_{\text{tHb}} = 47.5\ \mu\text{M}$. (d) NIRO-200NX large single-use vs. OxiplexTS in phantom 2. The sensor was repositioned after $c_{\text{tHb}} = 30\ \mu\text{M}$. There were two consecutive deoxygenations at $c_{\text{tHb}} = 30\ \mu\text{M}$ and $75\ \mu\text{M}$.




Fig. 3. (a) INVOS adult vs. OxiplexTS for three different c_{tHb} measured in phantom 3. There were 3 deoxygenations with sensor repositioning in between at $c_{tHb} = 47.5 \mu M$ with the second one being performed with a new sensor. Data from [38] is marked with (old). (b) OxyPrem v1.3 vs. OxiplexTS in phantom 3. There were 2 deoxygenations at $c_{tHb} = 47.5 \mu M$ with sensor repositioning in between. Data from [38] is marked with (old).

Table 4. Coefficients for linear transformation of StO_2 [%] from any oximeter to the scale of oximeters investigated in [38]: $StO_{2,to} = a * StO_{2,from} + b$. Coefficients were determined at $c_{tHb} = 47.5 \mu M$. Data at $c_{tHb} = 45 \mu M$ reprinted from [38] are printed *italic*. Part 1.

to → from ↓	OxiplexTS		OxyPrem v1.3		INVOS adult		INVOS neonatal		SenSmart neonatal	
	b	a	b	a	b	a	b	a	b	a
Oxiplex TS	0.0	1.00	21.3	0.71	8.1	1.00	11.8	1.09	47.1	0.41
<i>OxyPrem v1.3</i>	-30.0	1.41	0.0	1.00	-21.9	1.41	-21.0	1.54	34.8	0.58
<i>INVOS adult</i>	-8.1	1.00	15.6	0.71	0.0	1.00	3.0	1.10	43.8	0.41
<i>INVOS neonatal</i>	-10.8	0.91	13.6	0.65	-2.7	0.91	0.0	1.00	42.7	0.38
<i>SenSmart neonatal</i>	-114.9	2.44	-60.1	1.73	-106.4	2.43	-113.8	2.67	0.0	1.00
FORE-SIGHT small	-93.1	2.14	-44.7	1.52	-84.7	2.13	-90.0	2.34	8.9	0.88
FORE-SIGHT small band	-91.9	2.08	-43.9	1.48	-83.5	2.08	-88.7	2.28	9.4	0.85
NIRO small	-54.9	1.67	-17.7	1.18	-46.7	1.67	-48.3	1.83	24.6	0.69
NIRO small RU	-63.5	1.76	-23.7	1.25	-55.2	1.75	-57.6	1.92	21.1	0.72
INVOS adult	-4.8	0.97	17.9	0.68	3.2	0.96	6.5	1.06	45.1	0.40
OxyPrem V1.3	-26.4	1.35	2.6	0.96	-18.2	1.35	-17.0	1.48	36.3	0.55
SenSmart adult	15.0	1.17	31.9	0.83	23.0	1.17	28.2	1.28	53.3	0.48
FORE-SIGHT adult	-76.5	1.91	-33.0	1.35	-68.2	1.90	-71.9	2.09	15.7	0.78
NIRO large	-35.6	1.33	-4.0	0.95	-27.4	1.33	-27.1	1.46	32.5	0.55
NIRO large RU	-42.7	1.44	-9.0	1.02	-34.5	1.44	-34.9	1.58	29.6	0.59

Table 5. Coefficients for linear transformation of StO_2 [%] from any oximeter to the scale of any other oximeter: $\text{StO}_{2,\text{to}} = a * \text{StO}_{2,\text{from}} + b$. Coefficients were determined at $c_{\text{tHb}} = 47.5 \mu\text{M}$. Names of oximeters investigated in [38] at $c_{\text{tHb}} = 45 \mu\text{M}$ are printed *italic*. Part 2.

to → from ↓	FORE-SIGHT small		FORE-SIGHT small band		NIRO small		NIRO small RU		INVOS adult	
	b	a	b	a	b	a	b	a	b	a
OxiplexTS	43.5	0.47	44.1	0.48	32.9	0.60	36.1	0.57	5.0	1.04
<i>OxyPrem v1.3</i>	29.5	0.66	29.7	0.68	14.9	0.84	19.0	0.80	-26.1	1.46
<i>INVOS adult</i>	39.7	0.47	40.2	0.48	28.0	0.60	31.5	0.57	-3.4	1.04
<i>INVOS neonatal</i>	38.5	0.43	38.9	0.44	26.4	0.55	30.0	0.52	-6.2	0.95
<i>SenSmart neonatal</i>	-10.2	1.14	-11.0	1.17	-35.9	1.46	-29.2	1.39	-113.9	2.52
FORE-SIGHT small	0.0	1.00	-0.6	1.03	-22.8	1.28	-16.8	1.22	-91.3	2.21
FORE-SIGHT small band	0.5	0.97	0.0	1.00	-22.1	1.25	-16.2	1.19	-90.2	2.16
NIRO small	17.8	0.78	17.7	0.80	0.0	1.00	4.9	0.95	-51.9	1.73
NIRO small RU	13.8	0.82	13.7	0.84	-5.1	1.05	0.0	1.00	-60.7	1.82
INVOS adult	41.3	0.45	41.8	0.46	30.0	0.58	33.4	0.55	0.0	1.00
OxyPrem V1.3	31.2	0.63	31.5	0.65	17.1	0.81	21.1	0.77	-22.3	1.40
SenSmart adult	50.5	0.55	51.3	0.56	41.9	0.70	44.7	0.67	20.6	1.21
FORE-SIGHT adult	7.7	0.89	7.4	0.92	-12.9	1.14	-7.4	1.09	-74.2	1.98
NIRO large	26.9	0.62	27.0	0.64	11.6	0.80	15.8	0.76	-31.9	1.38
NIRO large RU	23.6	0.67	23.6	0.69	7.3	0.86	11.8	0.82	-39.2	1.49

Table 6. Coefficients for linear transformation of StO_2 [%] from any oximeter to the scale of any other oximeter: $\text{StO}_{2,\text{to}} = a * \text{StO}_{2,\text{from}} + b$. Coefficients were determined at $c_{\text{tHb}} = 47.5 \mu\text{M}$. Names of oximeters investigated in [38] at $c_{\text{tHb}} = 45 \mu\text{M}$ are printed *italic*. Part 3.

to → from ↓	OxyPrem v1.3		SenSmart adult		FORE-SIGHT adult		NIRO large		NIRO large RU	
	b	a	b	a	b	a	b	a	b	a
OxiplexTS	19.5	0.74	-12.8	0.85	40.1	0.52	26.7	0.75	29.6	0.69
<i>OxyPrem v1.3</i>	-2.7	1.04	-38.4	1.20	24.4	0.74	4.2	1.06	8.8	0.98
<i>INVOS adult</i>	13.5	0.74	-19.7	0.86	35.9	0.53	20.6	0.75	24.0	0.70
<i>INVOS neonatal</i>	11.5	0.68	-22.0	0.78	34.4	0.48	18.6	0.69	22.1	0.63
<i>SenSmart neonatal</i>	-65.5	1.81	-110.8	2.08	-20.1	1.28	-59.4	1.83	-50.1	1.69
FORE-SIGHT small	-49.4	1.58	-92.2	1.82	-8.7	1.12	-43.1	1.60	-34.9	1.48
FORE-SIGHT small band	-48.5	1.54	-91.2	1.78	-8.1	1.09	-42.2	1.56	-34.1	1.45
NIRO small	-21.2	1.24	-59.7	1.43	11.3	0.88	-14.5	1.25	-8.5	1.16
NIRO small RU	-27.5	1.30	-67.0	1.50	6.8	0.92	-20.9	1.32	-14.4	1.22
INVOS adult	16.0	0.72	-16.9	0.82	37.6	0.51	23.1	0.72	26.3	0.67
OxyPrem V1.3	0.0	1.00	-35.3	1.15	26.3	0.71	6.9	1.01	11.3	0.94
SenSmart adult	30.7	0.87	0.0	1.00	48.0	0.61	38.0	0.88	40.0	0.81
FORE-SIGHT adult	-37.1	1.41	-78.1	1.63	0.0	1.00	-30.7	1.43	-23.5	1.32
NIRO large	-6.8	0.99	-43.2	1.14	21.4	0.70	0.0	1.00	4.9	0.93
NIRO large RU	-12.1	1.07	-49.2	1.23	17.7	0.76	-5.3	1.08	0.0	1.00

4. Discussion

This study is an extension to the study in [38] with the aim to provide clinical researchers and clinicians a means to translate neonatal cerebral StO_2 acquired by their oximeter to the results reported by others, using other oximeters. As we used the same methods, we do not repeat the general discussion about aspects of the phantom but focus on new findings.

4.1. Observations

Relations between all oximeters investigated and OxiplexTS were linear except for these exceptions: (1) Fig. 1(a), 1(b) and 1(c) reveal that the FORE-SIGHT Elite oximeter is less sensitive to oxygenation changes for $\text{StO}_2 > 85\%$. Here the relation is non-linear. This does not affect the coefficients because we set the fitting range $16 \leq \text{StO}_2 \leq 85\%$ to circumvent this problem. The slightly smaller fitting range compared to [38] does not affect the results because there was little noise and all relations were linear with high R^2 (Table 2) except for SenSmart adult (see (3) below). (2) In Fig. 2(b) the $c_{\text{tHb}} = 30 \mu\text{M}$ curve of NIRO small is bent at $\text{StO}_2 > 85\%$, which has no influence on the results. This coincides with a manual pH measurement. It is probable that the hand-held pH-probe accidentally partially obstructed the light path of the sensor. (3) SenSmart adult to OxiplexTS relation is non-linear (Fig. 1(d)). We have observed this already in a previous study [42]. Although the linear fits (Table 2) here depend on the StO_2 range of fitting, we consider the deviation from the fit lines acceptable within the fit range, but extrapolation to higher or lower StO_2 is inaccurate.

Not all manufacturers published their algorithms and therefore it is unknown why instruments over-/underestimate StO_2 . One reason may be unaccounted absorbers such as water (H_2O) present in the phantom, particularly during low c_{tHb} [39]. The phantom contains $\sim 98\%$ H_2O and is similar to brain tissue of neonates with up to 95% H_2O [50]. This is discussed in detail in [38].

We observed a decrease in phantom μ'_s over time. The μ'_s depends on the number and size of sub-micrometer lipid droplets of the IL emulsion and their relative refractive index. We assume the decrease in μ'_s to be caused by confluence of these droplets into larger droplets. The StO_2 errors due to the variation in μ'_s were small. In repeated deoxygenations, there is no relevant deviation visible (Fig. 1(a), 1(c), 1(d), 2(b), 2(d), 3(a), and 3(b)) except for the small one due to sensor repositioning ($< 2\%$, see Sec. 4.3). We therefore conclude that the phantom experiments may last up to 4.5 hours, as long as the decrease in μ'_s is in the range of this experiment ($\leq 16\%$).

We cannot tell why an error was reported after removal and reattachment of the adhesive sensor (FORE-SIGHT small). The values of this sensor agreed well to FORE-SIGHT small band beforehand (see Table 2), as expected since both sensors apparently are similar.

We do not report results for FORE-SIGHT medium, because its response at $c_{\text{tHb}} = 75 \mu\text{M}$ compared to the response at the lower c_{tHb} was too different from the patterns observed with the large and both small sensors of the same oximeter. FORE-SIGHT medium data from another, unpublished experiment also showed different patterns for all three c_{tHb} . Simultaneous StO_2 readings of NIRO small RU agreed well with another unpublished experiment. Thus, we exclude phantom 2 as reason for this deviation. We also exclude inadequate sensor placement because the sensor was repositioned several times. Probably the specific sensor used was simply defective and consequently we do not report results for FORE-SIGHT medium.

The non-linear behavior for the SenSmart adult was also observed in a previous phantom study [42], but with remarkably stronger curvature in the present study (Fig. 1(d)). In phantom 3 at $c_{\text{tHb}} = 47.5 \mu\text{M}$, one measurement with a brand new SenSmart adult sensor agrees well with the other measurements (Fig. 1(d)). We therefore exclude a sensor defect. SenSmart adult was used in two different phantoms and the sensor was repositioned several times, hence we exclude inadequate sensor placement. Simultaneous readings of other oximeters in both phantoms were inconspicuous. We consequently exclude the phantoms 2 and 3 as reasons causing this observation. One difference between the present study and [42] are the different optical properties

for the windows. Both sets of windows were made from the same ingredients, but different quantities. At 692 and 834 nm, respectively, the windows of current study and [38] had μ'_a of 0.059 and 0.057 [1/cm], and μ'_s of 5.0 and 4.4 [1/cm], while the windows of [42] had μ'_a of 0.10 and 0.11 [1/cm], and μ'_s of 9.6 and 8.3 [1/cm]. Accordingly, we speculate that in contrast to other oximeters, SenSmart adult may be sensitive to optical properties of the superficial layer, although it is only 2.5 mm thick. We decided to report results of SenSmart adult for comparison to [42], but this issue needs to be further investigated.

4.2. Comparison to *in vivo* findings

Hessel et al. compared INVOS neonatal and an older version of FORE-SIGHT small in neonates directly after birth [4]. They simultaneously recorded with sensors placed on separate hemispheres of the head. Their StO₂ data is plotted in Fig. 4 with the fit line we obtained *in vitro* $c_{tHb} = 47.5 \mu\text{M}$ (table 4). The line lies within the point clouds, demonstrating that our *in vitro* agree reasonably well with *in vivo* data [4]. Please note that the old FORE-SIGHT instrument had 4 wavelengths generated by laser diodes, while the new FORE-SIGHT Elite used here employed 5 wavelengths generated by LEDs. Thus the instruments are quite different. We did not find any literature comparing the two versions of the instrument with neonatal sensors. Assuming that the older and the new version of the instrument provide similar StO₂ values, Figure 4 would indicate reasonable agreement between *in vivo* and *in vitro* measurements which adds plausibility to our *in vitro* evaluation.




Fig. 4. *In vivo* comparison [4] of simultaneous StO₂ recordings in neonates with the old FORE-SIGHT instrument (FS1) with small sensor and INVOS neonatal as scatter density plot (gray dots, darker means higher occurrence). The red line corresponds to the deming regression [51] of this data ($y = 1.98x - 73.7$). This agrees reasonably with the relation (dashed green line) we obtained in this *in vitro* phantom study with the new FORE-SIGHT Elite instrument (FS Elite) at $c_{tHb} = 47.5 \mu\text{M}$ ($y = 2.34x - 90$, see Table 4).

4.3. Reproducibility of the method

In our study, sensors were un-mounted and re-mounted between repeated measurements at the same c_{tHb} (Fig. 1(a), 2(b), 1(d), 3(a), and 3(b)). The sensor repositioning causes slight variation in sensor location and pressure applied to the window. Repositioning errors were $< 2\%$ StO_2 , which we consider acceptable.

For INVOS adult and SenSmart adult we attached a brand-new sensor for one deoxygenation at $c_{\text{tHb}} = 47.5\ \mu\text{M}$ in phantom 3. Both showed exactly the same behavior as the old sensors (Fig. 1(d) and 3(a)), suggesting negligible variation between individual sensors.

INVOS adult and OxyPrem v1.3 oximeters were investigated in [38] and in the present study (Fig. 3(a) and 3(b)) enabling to determine the repeatability. Within the fitting range, differences were $< 3\%$. Table 4 shows that linear transformation from present results (in **rows**) to those of [38] (in **columns**) is $y = 0.96 * x + 3.2$ for INVOS adult and $y = 0.96 * x + 2.6$ for OxyPrem v1.3 and thus close to the perfect relationship of $y = 1 * x + 0$. Table 3 further shows that uncertainty range due to changes in c_{tHb} is slightly increased compared to [38] for OxyPrem v1.3 at the hypoxic threshold and for INVOS adult at the hyperoxic threshold. This increase is most likely caused by the sensor re-mounting between the deoxygenations at $c_{\text{tHb}} = 47.5\ \mu\text{M}$ which resulted in a slight shift of values (Fig. 3(b)). Compared to [38], estimation of SafeBoosC action thresholds only deviates by 1 % at the hypoxic threshold for INVOS adult and by 1 % at hyperoxic threshold for OxyPrem v1.3. We conclude that the phantom showed good reproducibility.

4.4. Implications of other haemoglobin species

The tested oximeters do not account for dyshaemoglobins, e.g. carboxy- and methaemoglobin, or fetal Hb, so their presence could be a source of error depending on their concentration and extinction coefficients [31]. We used adult donor blood for the phantom and hence fetal Hb was not present. Fetal Hb has nearly the same absorption spectra as adult Hb [52]. So, even in neonatal populations when fetal Hb may be dominant, this would at most introduce a negligible error [53]. Methaemoglobin has extinction coefficients in the near-infrared range that are close to those of oxygenated haemoglobin (O_2Hb), while the carboxyhaemoglobin extinction coefficient is negligible [52]. Normally, methaemoglobin is $< 2\%$ of all Hb. Methaemoglobin was not determined in the donor blood, but only increases slightly during storage [54]. No known agents for the formation of methaemoglobin were used during the experiment and hence, we do not consider methaemoglobin a likely source of error.

5. Conclusion

Extending our previous findings to further oximeter brands, we confirmed that different brands of oximeters provide different tissue oxygen haemoglobin saturation (StO_2) readings in a phantom designed to mimic the head of preterm infants. Preliminary clinical data are in agreement with the phantom results. We provided linear equations which translate data from one oximeter to another at a total haemoglobin concentration (c_{tHb}) typical for preterm infants. Accordingly, intervention thresholds have to be set specifically for each brand and sensor type. Additionally, these thresholds are subject to uncertainties arising from device-specific dependence of the StO_2 readings on the c_{tHb} . Previous results were reproduced with $< 3\%$ deviation. Reproducibility of hypoxic and hyperoxic thresholds was $\approx 1\%$, and variation due to sensor removal and replacement was $< 2\%$. This demonstrates good reproducibility of the method.

Funding

Danish Council for Strategic Research (grant number 00603-00482B); the Nano-Tera projects ObeSense, ParaTex and NewbornCare; Swiss National Science Foundation (project 159490).

Acknowledgments

The authors would like to thank Trine W. Hessel and her co-authors [4] who supported this study by sharing their raw data. We further want to thank CSMED and Hamamatsu who provided us with their instruments and sensors for the purpose of this experiment.

Disclosures

The authors declare that there are no conflicts of interest related to this article.